

**Test Report No. 7191256991-EEC21/01B-WBH**  
dated 11 May 2021



PSB Singapore

**Add value.  
Inspire trust.**

**Note:** This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

**SUBJECT**

Testing of Gloves

**CLIENT**

Shijiazhuang Hongray Group Co.,Ltd.  
South Tongda Rd., East Dist.,  
Jinzhou City, Hebei,  
052260, China

**SAMPLE SUBMISSION DATE/ TEST DATE**

29 Mar 2021/ 01 Apr 2021 to 06 May 2021

**DESCRIPTION OF SAMPLES**

S/N	Product Description	Brand/ Model	Colour	Lot No.	Size	Sample Received (pieces)	Manufacturer
1	Disposable Nitrile Gloves	Hongray	Blue	- (See Remark 1)	S	400	Shijiazhuang Hongray Group Co.,Ltd.
2					M	400	
3					L	400	
4					XL	400	

**METHOD OF TEST:**

1. EN 455-1:2020 Medical gloves for single use  
Part 1: Requirements and testing for freedom from holes
2. EN 455-2:2015 Medical gloves for single use  
Part 2: Requirements and testing for physical properties
3. EN 455-3:2015 Medical glove for single use  
Part 3: Requirements and testing for biological evaluation



Laboratory:  
TÜV SÜD PSB Pte. Ltd.  
15 International Business Park  
TÜV SÜD @ IBP  
Singapore 609937

Phone : +65-6778 7777  
E-mail: info.sg@tuvsud.com  
<https://www.tuvsud.com/sg>  
Co. Reg : 199002667R

Regional Head Office:  
TÜV SÜD Asia Pacific Pte. Ltd.  
15 International Business Park  
TÜV SÜD @ IBP  
Singapore 609937  
**TUV®**

**Test Report No. 7191256991-EEC21/01B-WBH**  
dated 11 May 2021



PSB Singapore

**RESULTS:**

Sample: Disposable Nitrile Gloves, Hongray, Blue

Table 1: Results for EN 455-1:2020

Clause	Tests	Size	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4 5	Freedom from holes	M	Shall not leak	10	315	2	Passed

Table 2: Results for EN 455-2:2015 Clauses 4-5

Clause	Tests	Size	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	Dimensions a) Length (mm)	S	≥ 240	13	250	Passed
		M		13	249	Passed
		L		13	246	Passed
		XL		13	248	Passed
	b) Width (mm)	S	80 ± 10	13	84	Passed
		M	95 ± 10	13	96	Passed
		L	110 ± 10	13	115	Passed
		XL	≥ 110	13	114	Passed
5	Strength a) Force at break (N)	S	For nitrile examination gloves: ≥ 6.0	13	7.5	Passed
		M		13	8.1	Passed
		L		13	8.3	Passed
		XL		13	7.5	Passed
	b) Force at break after challenge testing (N) 7 days at (70±2)°C	S	For nitrile examination gloves: ≥ 6.0	13	7.6	Passed
		M		13	8.1	Passed
		L		13	8.0	Passed
		XL		13	7.7	Passed

Table 3: Results for EN 455-2:2015 Clause 7

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed

**Test Report No. 7191256991-EEC21/01B-WBH**  
dated 11 May 2021



PSB Singapore

**RESULTS (cont'd):**

Sample: Disposable Nitrile Gloves, Hongray, Blue

Table 4: Results for EN 455-3:2015 Clauses 4.2-4.5


Clause	Tests	Requirements	Results / Remarks	Inferred results	
4.2	Chemicals	Gloves shall not be dressed with talcum powder (magnesium silicate).	Glove is not dressed with talcum powder, based on client's declaration letter	Passed	
		Other chemicals	Manufacturer shall disclose upon request a list of chemical ingredients	NA	
4.3 5.1	Endotoxins	< 20 EU/pair for gloves labelled with 'low endotoxin content'.	Not labelled with 'low endotoxin content'	NA	
4.4 5.2	Powder-free gloves	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove.	S	1.24 mg per glove	Passed
			M	0.91 mg per glove	Passed
			L	0.83 mg per glove	Passed
			XL	0.27 mg per glove	Passed
4.5 5.3	Proteins, leachable	The manufacturer shall strive to minimize the leachable protein level for gloves containing natural rubber latex.	Not natural rubber latex glove	NA	


Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex;	NA
		The labelling shall include the following or equivalent warning statement together with the symbol: '(Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses';	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free;	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: 'CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions';	NA
		d) for any medical glove containing natural rubber latex the product labelling shall not include: - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unjustified indication of the presence of allergens;	NA
e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 5.3 shall be given.	NA		
Inferred results			Passed

**REMARKS:**

1. Lot No. was not provided by client.
2. Labelling requirements were assessed based on submitted packaging artwork by client.
3. NA: Not applicable for the submitted sample.
4. The colour as "Blue" was declared by client.

  
\_\_\_\_\_  
Yeo Poh Kwang  
Associate Engineer

  
\_\_\_\_\_  
Wong Bee Hui  
Product Manager  
Medical Health Services (NAM)

**APPENDIX:**



Photo 1: Disposable Nitrile Gloves, Hongray, Blue

SUD

**Test Report No. 7191256991-EEC21/01B-WBH**  
**dated 11 May 2021**



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, 15 International Business Park TÜV SÜD @ IBP Singapore 609937.
6. The tests carried out by TÜV SÜD PSB and this report are subject to TÜV SÜD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

Effective 26 January 2021

