

EU Declaration of Conformity

Manufacturer

Shijiazhuang Hongray Group Co., Ltd.

South Tongda Rd., East Dist. Jinzhou,052260

Hebei, China.

Tel: +86-311-83610904

Fax: +86-311-83610904

SRN:CN-MF-000024227

European Representative

Caretechion GmbH

Niederrheinstr.71, 40474 DüsseldorfGermany

DIMDI Code: DE/0000048026

Tel/Fax: 0211 30033 6681

Email: info@caretechion.de

SRN: DE-AR-000005946

The undersigned Shijiazhuang Hongray Group Co., Ltd., as manufacturer,

Setino Hungary Kft as the importer and distributor of:

NITRILE EXAMINATION AND PROTECTIVE GLOVES (WITH SETINO ® TRADEMARK)

Intended Purpose: A patient examination glove is a disposable device intended for medical purposes that is worn on the examiners hand or finger to prevent contamination between patient and examiner. Examination glove is intended for medical activities except surgery.

BASIC UDI-DI: 9487944

Reference number	Product name
NPF3001B	Nitrile Examination and Protective Gloves,black,XS
NPF3002B	Nitrile Examination and Protective Gloves, black, S
NPF3003B	Nitrile Examination and Protective Gloves, black, M
NPF3004B	Nitrile Examination and Protective Gloves, black, L
NPF3005B	Nitrile Examination and Protective Gloves, black, XL

Declares that

the product complies with the terms of EU Regulation 2017/745 (MDR) and conform to the general safety and performance requirements (Annex I) as well as to the following applicable technical norms mentioned in the technical files:

EN455-1:2000, EN455-2:2015, EN455-3:2015, EN455-4:2009

Based on Annex VIII of MDR 2017/745 EU, the medical device has been classified as belonging to I class.

 $C \in$

The company has certified its own Quality and Management System in conformity to the norm EN ISO 9001:2015(Certificate no. **011001732303** given on 13/02/2023 by TUV Rheinland).

This EU declaration of conformity has been drafted based on annex IV of MDR 2017/745 under sole responsibility of

Shijiazhuang Hongray Group Co., Ltd.

South Tonda Rd., East Dist. Jinzhou, 052260 Hebei, China

Place and date

QA manager:Wu Min